



Traceable Certificate Number: 2818903
Contractor: KOENIG SCALE
 4779 E MARGARET DR
 TERRE HAUTE, IN 47803-9303

Purchase Order Number: 20505
Client: KOENIG SCALE
 4779 E MARGARET DR
 TERRE HAUTE, IN 47803-9303

Date Received: 19 Oct 2018
Date Calibrated: 23 Oct 2018
Recall Date: 23 Oct 2019
Temperature Range: 21.62 °C
Pressure Range: 740.01 mmHg
Relative Humidity Range: 46 %
Air Density Range: 1.1611 mg/cm³
NIST Certificate Number: 684/286541-15 & 684/290551-18

* Although this test weight has been tested for the tolerance of the class for which it was submitted, it does not meet the finish requirements for that class

Although there are two NIST numbers, one or both may apply

Calibrated By: 05
Procedure: Inter-comparison Method (WI05-0095)
Condition of Weights: Acceptable for Calibration
Description of Weights: 50 lb Satin Finish Grip Handle Weight, ASTM Class 1, S/N 14ZX

Nominal Value	ID or S/N	As Found			As Left			Unc. (mg)	k	MPE* (mg)	Balance Used	Standard Set Used	Assumed Density (g/cm ³)
		Conv. Mass	Conv. Mass Corr (mg)	MPE Pass	Conv. Mass	Conv. Mass Corr (mg)	MPE Pass						
50 lb	14ZX *	49.999990		-5 Y	49.999990		-5 Y	13	2	57	1771Q	N568Q	7.84

This report contains data not covered by the NVLAP Accreditation if the box is checked.

Check with your local state agency for certification of compliance on Legal for Trade items. *The weight accuracy class is referenced in the Description of Weights. Unless otherwise noted, the weights calibrated meet the requirements of the accuracy class. The Uncertainty of Measurement is included in the determination of Maximum Permissible Error (MPE) Pass/Fail Criteria. The specifications for Maximum Permissible Error (MPE) can be found in NIST Handbook

Prepared By: 105-1 (1990), ASTM E617-13 or OIML R111-1 (2004), manufacturer specifications or customer specifications.

Rice Lake Weighing Systems

230 West Coleman Street, Rice Lake, WI 54868 • USA • PN 64787 • 3/18

TEL: 715-234-9171 • FAX: 715-234-6967 • www.ricelake.com

Definitions: <http://certs.ricelake.com/certs/DefinitionsV2.docx>

Dated 23 Oct 2018

Dan Demers
 Dan Demers, Metrologist



The Uncertainty assigned to the Conventional Mass values are the result of the root-sum-square of the type A and type B components, calculated in accordance with NIST SOP 29 and ISO GUM, with a coverage factor (k), to express the expanded uncertainty with an approximate 95.45 % confidence level. This Report is not to be used to claim product certification, approval, or endorsement by NVLAP, NIST, A2LA or any agency of the U.S. Government. This document shall not be reproduced, except in full, without the written approval of Rice Lake Weighing Systems' Metrology Laboratory.



Chris D Bradford 11/26/2018