



Calibration Certificate

Certificate Number: 20-198

Calibration Date: 07/21/2020

Calibration Due Date: 07/21/2021

Tested Item(s): 6 Metric Test Kits

Issued To: Koenig Scale Company Inc
4779 E. Margaret Dr.
Terre Haute IN, 47803

POC: Jim Maxwell
Phone: 812-877-6121

Date Received: 07/15/2020

Artifact(s) Description

Test Item(s): 6 Metric Test Kits

Serial No: KS-100, 2STG, 8M, 4M, 2M, 1M

Manufacture: Rice Lake

Material: Stainless Steel

Class Specification: NIST HB 105-1 (1990), Class F

Condition: Fair

Calibration Information

Metrologist: Howard Wickersham

Procedure: NISTIR 6969, SOP 8

Temperature: 21.2 °C

Pressure: 740.4 mm Hg

Humidity: 51.7 % RH

Summary Table

Nominal Mass	Serial No. / ID	Manufacture	Correct	Adjusted	Condemned	Confiscated	Total
Various	KS-100	Rice Lake	4	0	0	0	4
Various	2STG	Rice Lake	1	0	0	0	1
Various	8M	Rice Lake	10	0	0	0	10
Various	4M	Rice Lake	14	0	0	0	14
Various	2M	Rice Lake	13	0	0	0	13
Various	1M	Rice Lake	18	0	0	0	18
Grand Total			60	0	0	0	60

Calibration Results

Nominal Mass	Serial No. / ID	Manufacture	Conventional Mass Correction		U ± (mg)	k factor	Degrees of Freedom	NIST HB 105-1 (1990), Class F MPE ± (mg)	Assumed Density (g/cm ³)
			As Found (mg)	As Left (mg)					
5 kg	2STG	Rice Lake	208	208	68	2.32	9	500	7.95
5 kg	KS-100-A	Rice Lake	162	162	68	2.32	9	500	7.95
5 kg	KS-100-B	Rice Lake	153	153	68	2.32	9	500	7.95
5 kg	KS-100-C	Rice Lake	155	155	68	2.32	9	500	7.95
5 kg	KS-100-D	Rice Lake	142	142	68	2.32	9	500	7.95
2 kg	2M	Rice Lake	32	32	24	2.025	54	200	7.95

This certificate shall not be reproduced except in full, without written approval by the laboratory.



Calibration Certificate

Certificate Number: 20-198

Calibration Date: 07/21/2020

Calibration Due Date: 07/21/2021

Tested Item(s): 6 Metric Test Kits

2 kg	8M	Rice Lake	33	33	24	2.025	54	200	7.95
2 kg	1M	Rice Lake	80	80	24	2.025	54	200	7.95
2 kg	4M	Rice Lake	42	42	24	2.025	54	200	7.95
1 kg	1M	Rice Lake	17	17	13	2	7783	100	7.95
1 kg	4M	Rice Lake	29	29	13	2	7783	100	7.95
1 kg	8M	Rice Lake	26	26	13	2	7783	100	7.95
1 kg	2M	Rice Lake	6	6	13	2	7783	100	7.95
500 g	1M	Rice Lake	46.6	46.6	9	2.078	66	70	7.95
500 g	2M	Rice Lake	34.6	34.6	9	2.078	66	70	7.95
500 g	4M	Rice Lake	25.6	25.6	9	2.078	66	70	7.95
500 g	8M	Rice Lake	10.6	10.6	9	2.078	66	70	7.95
200 g	1M-1	Rice Lake	-0.3	-0.3	4.8	2	10732	40	7.95
200 g	1M-2	Rice Lake	1.6	1.6	4.8	2	10732	40	7.95
200 g	2M-Dot	Rice Lake	7.4	7.4	4.8	2	10732	40	7.95
200 g	2M-No Dot	Rice Lake	7.6	7.6	4.8	2	10732	40	7.95
200 g	4M-1	Rice Lake	8.9	8.9	4.8	2	10732	40	7.95
200 g	4M-2	Rice Lake	13	13	4.8	2	10732	40	7.95
200 g	8M-Dot	Rice Lake	16.4	16.4	4.8	2	10732	40	7.95
200 g	8M-No Dot	Rice Lake	7.1	7.1	4.8	2	10732	40	7.95
100 g	1M	Rice Lake	4.9	4.9	2.4	2.025	85	20	7.95
100 g	4M	Rice Lake	6.3	6.3	2.4	2.025	85	20	7.95
100 g	8M	Rice Lake	7	7	2.4	2.025	85	20	7.95
50 g	1M	Rice Lake	3.2	3.2	1.2	2.025	92	10	7.95
50 g	2M	Rice Lake	2.5	2.5	1.2	2.025	92	10	7.95
50 g	4M	Rice Lake	5.2	5.2	1.2	2.025	92	10	7.95
50 g	8M	Rice Lake	6	6	1.2	2.025	92	10	7.95
20 g	1M-Dot	Rice Lake	1.6	1.6	0.48	2.025	86	4	7.95
20 g	1M-No Dot	Rice Lake	2.38	2.38	0.48	2.025	86	4	7.95
20 g	2M-Dot	Rice Lake	0.56	0.56	0.48	2.025	86	4	7.95
20 g	2M-No Dot	Rice Lake	1.6	1.6	0.48	2.025	86	4	7.95
20 g	4M-Dot	Rice Lake	2.81	2.81	0.48	2.025	86	4	7.95
20 g	4M-No Dot	Rice Lake	2.26	2.26	0.48	2.025	86	4	7.95
20 g	8M	Rice Lake	1.27	1.27	0.48	2.025	86	4	7.95
10 g	1M	Rice Lake	0.65	0.65	0.27	2.025	87	2	7.95
10 g	2M	Rice Lake	0.52	0.52	0.27	2.025	87	2	7.95
10 g	4M	Rice Lake	0.56	0.56	0.27	2.025	87	2	7.95
10 g	8M	Rice Lake	1.02	1.02	0.27	2.025	87	2	7.95
5 g	1M	Rice Lake	0.64	0.64	0.2	2.05	58	1.5	7.95
5 g	2M	Rice Lake	-0.13	-0.13	0.2	2.05	58	1.5	7.95
5 g	4M	Rice Lake	0.65	0.65	0.2	2.05	58	1.5	7.95
5 g	8M	Rice Lake	0.61	0.61	0.2	2.05	58	1.5	7.95
2 g	1M-Dot	Rice Lake	0.39	0.39	0.15	2.025	58	1.1	7.95
2 g	1M-No Dot	Rice Lake	0.73	0.73	0.15	2.025	58	1.1	7.95
2 g	2M-Dot	Rice Lake	0.45	0.45	0.15	2.025	58	1.1	7.95

This certificate shall not be reproduced except in full, without written approval by the laboratory.



Calibration Certificate

Certificate Number: 20-198

Calibration Date: 07/21/2020

Calibration Due Date: 07/21/2021

Tested Item(s): 6 Metric Test Kits

2 g	2M-No Dot	Rice Lake	0.56	0.56	0.15	2.025	58	1.1	7.95
2 g	4M-Dot	Rice Lake	0.5	0.5	0.15	2.025	58	1.1	7.95
2 g	4M-No Dot	Rice Lake	0.56	0.56	0.15	2.025	58	1.1	7.95
1 g	1M	Rice Lake	0.43	0.43	0.12	2.32	9	0.9	7.95
1 g	2M	Rice Lake	0.16	0.16	0.12	2.32	9	0.9	7.95
1 g	4M	Rice Lake	0.27	0.27	0.12	2.32	9	0.9	7.95
500 mg	1M	Rice Lake	0.33	0.33	0.11	2.65	5	0.72	7.95
200 mg	1M-Dot	Rice Lake	0.22	0.22	0.07	2.25	11	0.54	7.95
200 mg	1M-No Dot	Rice Lake	0.16	0.16	0.07	2.25	11	0.54	7.95
5 mg	1M	Rice Lake	0.036	0.036	0.026	2.65	5	0.17	7.95

Traceability Statement

The artifact(s) described in this calibration certificate have been compared to the Standards of the State of Indiana. The Standards of the State of Indiana are traceable to the National Institute of Standards and Technology (NIST) and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The International System of Units (SI) for mass is the kilogram (kg) (see Conversion Factors below). The certificate number for this calibration is the only unique number to be used in referencing measurement traceability for the artifact(s) described in this calibration certificate.

Uncertainty Statement

The combined standard uncertainty includes uncertainties reported for the standard, uncertainties associated with the measurement process, uncertainties for any observed deviations from reference values which are less than surveillance limits (previous similar determinations have demonstrated that the maximum permissible errors are sufficiently large that buoyancy corrections are not usually significant [i.e., corrections & their uncertainty will not change the last decimal place of the calibration value or uncertainty (with uncertainty rounded to 2 significant digits)]. The combined standard uncertainty is multiplied by a coverage factor, k , to give the expanded uncertainty, which defines an interval with an approximate 95 % level of confidence. The expanded uncertainty presented in this calibration certificate is consistent with NIST Technical Note 1297. Surface Roughness and Magnetic testing has not been performed, therefore, there are no components for the effects of either in the uncertainty budget.

Pertinent Information

- In-accordance-with ISO/IEC FDIS 17025, General Requirements for the Competence of Testing and Calibration Laboratories, paragraph 5.10.4.4 'A calibration certificate (or calibration label) shall not contain any recommendation on the calibration interval except where this has been agreed with the client. This requirement may be superseded by legal regulations.'
- In-accordance-with Indiana Code (IC) 24-6-3-2, Division of weights and measures; powers and duties, Section 2(b), '...The division, or inspectors at the divisions direction, shall correct the standards of the several cities and counties, and as often as once in two (2) years compare the same with those in the division's possession, and where not otherwise provided by law the division shall have the general supervision of the weights, measures, and measuring and weighing devices in use in Indiana...'

This certificate shall not be reproduced except in full, without written approval by the laboratory.



Calibration Certificate

Certificate Number: 20-198

Calibration Date: 07/21/2020

Calibration Due Date: 07/21/2021

Tested Item(s): 6 Metric Test Kits

- The artifact(s) listed above have been found and/or left within the maximum permissible error for the specification stated above, except as noted. An artifact is considered in-compliance when the correction plus the measurement uncertainty is equal to or less than the maximum permissible error. **BOLD** print indicates an out-of-compliance reading.
- All corrections stated in this calibration certificate correlate to a "Conventional Mass" (CM), also known as 'apparent mass', scale verses 8.0 g/cm³ reference mass density and an air density of 1.2 mg/cm³ at 20 °C.
- The results listed in this calibration certificate relate only to the artifacts described and extent of calibrations performed.

Conversion Factors

From NIST Special Publication 811, *Guide for the Use of the International System of Units (SI)*

Factors in **boldface** are exact

To convert from	to	multiply by
carat, metric	to kilogram (kg)	2.0 E-04
grain (gr)	to kilogram (kg)	6.479 891 E-05
ounce (avoirdupois) (oz)	to kilogram (kg)	2.834 952 E-02
ounce (troy or apothecary) (oz)	to kilogram (kg)	3.110 348 E-02
dram (apothecary) (dr)	to kilogram (kg)	3.8879346 E-03
scruple (apothecary) (s)	to kilogram (kg)	1.2959782 E-03
pennyweight (dwt)	to kilogram (kg)	1.555 174 E-03
pound (avoirdupois) (lb)	to kilogram (kg)	4.535 923 7 E-01

I declare or certify under penalty of perjury under the laws of the State of Indiana that the foregoing is true and correct:

Signed on this 21st day of July, 2020 in the city of Indianapolis, Marion County, Indiana

Signature: Howard Wickersham

Howard Wickersham, State Metrologist, Technical Manager