



CALIBRATION CERTIFICATE

Cast Iron Weight, NIST Class F
Contents: (21) 1000 lb, (16) 500 lb, (6) 250 kg

Customer:

Koenig Scale Co, Inc.
4779 E. Margaret Dr
Terre Haute, IN 47803
Kevin Koenig
Quality Manager

Certificate Number: 2024-0874

Date Received: 7/15/2024

Date Calibrated: 7/16/2024

Condition of Artifacts: Acceptable

Page 1 of 2

Methods and Traceability:

The weights described above have been compared with the standards of the State of Ohio, which are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST). The weights are compliant with NIST Class F specifications and tolerances from NIST Handbook 105-1 (1990). The uncertainty of the measurement is considered when making this statement of compliance. Surface finish, density and magnetism were not evaluated. It is the end user's responsibility to verify that the weights meet the accuracy requirements outlined in NIST Handbook 44 (2020), Appendix A Fundamental Considerations, when using the weights for calibration of commercial (Legal for Trade) scales. ** The laboratory calculates measurement uncertainties in accordance with NIST Standard Operating Procedure 29 (2019); ISO Guide to the Expression of Uncertainty in Measurement (GUM) (JCGM 100:2008, GUM 1995 with minor corrections, 2008); and ISO/IEC GUIDE 98-3:2008, Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement (GUM:1995). The combined standard uncertainty is the result of the root-sum-square of the Type A and Type B components specified in the SOP listed below. The combined standard uncertainty is multiplied by an appropriate coverage factor (k) to provide an expanded uncertainty which defines an interval having a level of confidence of approximately 95 percent. The effects of magnetism have not been considered in the measurement uncertainty. ** Conventional mass correction values are reported. These are based on the mass in air with reference conditions of air density 0.0012 g/cm³, mass density of 8 g/cm³, and a reference temperature of 20 °C. The correction is the weight's error from its nominal value. As Received values indicate the error of the items as submitted, before adjustment.

The following Standard Operating Procedure (S.O.P.) was used:

- 8 NIST Recommended Standard Operating Procedure for Medium Accuracy Calibration of Mass Standards by Modified Substitution. (May 2019)

Notes:

[Handwritten signature of Keith Crider]

Weights & Measures Technologist Keith Crider

7/23/2024

Issue Date



NVLAP CODE: 200420

This report shall not be reproduced except in full, without written approval of the Ohio Weights & Measures Metrology Laboratory. Clients must not use this report to claim product endorsement by this laboratory, NVLAP or any agency of the U.S. government. The results of calibrations shown on this report apply only to the items listed and at the time of the test.



Serving Farmers and Protecting Consumers Since 1846

[Handwritten signature]

07/30/24

OHIO DEPARTMENT OF AGRICULTURE
DIVISION OF WEIGHTS AND MEASURES METROLOGY LABORATORY

SUBMITTED BY:	Koenig Scale Co, Inc.	CLASS:	NIST F	Page	2		
OHIO TEST NO.:	2024-0874	TEMP:	20.18	°C			
DATE:	7/16/2024	PRESS:	732.01	mm Hg			
DESCRIPTION:	Cast Iron Weight	HUMID:	51.76	% RH			
Weight Identification	Nominal Value	Units	Tolerance ± (mg)	As Received (mg)	As Left (mg)	Uncertainty ± (mg)	k value
1	1000	lb	45000	-1100	-1100	6200	2.011
2	1000	lb	45000	-4600	-4600	6200	2.011
3	1000	lb	45000	-900	-900	6200	2.011
4	1000	lb	45000	-11700	-11700	6200	2.011
5	1000	lb	45000	-25500	-25500	6200	2.011
6	1000	lb	45000	-8200	-8200	6200	2.011
7	1000	lb	45000	-1500	-1500	6200	2.011
8	1000	lb	45000	-19300	-19300	6200	2.011
9	1000	lb	45000	-21400	-21400	6200	2.011
10	1000	lb	45000	-24300	-24300	6200	2.011
11	1000	lb	45000	-5300	-5300	6200	2.011
12	1000	lb	45000	-27600	-27600	6200	2.011
13	1000	lb	45000	-16500	-16500	6200	2.011
14	1000	lb	45000	-13400	-13400	6200	2.011
15	1000	lb	45000	-20300	-20300	6200	2.011
16	1000	lb	45000	-15600	-15600	6200	2.011
17	1000	lb	45000	-27300	-27300	6200	2.011
18	1000	lb	45000	-19600	-19600	6200	2.011
19	1000	lb	45000	-17000	-17000	6200	2.011
20	1000	lb	45000	-19800	-19800	6200	2.011
G692	1000	lb	45000	-16200	-16200	6200	2.011
650L	500	lb	23000	-14900	-14900	3100	2.023
651L	500	lb	23000	-3500	-3500	3100	2.023
652L	500	lb	23000	4500	4500	3100	2.023
653L	500	lb	23000	-5600	-5600	3100	2.023
654L	500	lb	23000	-6600	-6600	3100	2.023
655L	500	lb	23000	-13300	-13300	3100	2.023
656L	500	lb	23000	-11900	-11900	3100	2.023
657L	500	lb	23000	11200	11200	3100	2.023
658L	500	lb	23000	2800	2800	3100	2.023
659L	500	lb	23000	-9200	-9200	3100	2.023
660L	500	lb	23000	11900	11900	3100	2.023
661L	500	lb	23000	-2200	-2200	3100	2.023
662L	500	lb	23000	8200	8200	3100	2.023
663L	500	lb	23000	0	0	3100	2.023
GL1	500	lb	23000	4900	4900	3100	2.023
GL2	500	lb	23000	7700	7700	3100	2.023
81	250000	g	25000	4800	4800	7800	2.011
82	250000	g	25000	12200	12200	7800	2.011
83	250000	g	25000	-9600	-9600	7800	2.011
84	250000	g	25000	6100	6100	7800	2.011
85	250000	g	25000	3200	3200	7800	2.011
86	250000	g	25000	-1300	-1300	7800	2.011

Chris D Bradford

07/30/24